Comparison of reabsorbable mini anchor *versus* non reabsorbable: histological evaluation in an experimental rabbit model

Alex Pontini¹
Ivan Munivrana¹
Maria Luisa Valente²
Franco Bassetto¹

¹ Department of Medico-Surgical Specialties, Hand Unit, Institute of Plastic Surgery, University of Padova, Italy

² Department of Diagnostic Medical Sciences, Section of Special Pathology, University of Padova, Italy

Corresponding author:

Alex Pontini
Department of Medico-Surgical Specialties, Hand Unit
Institute of Plastic Surgery, University of Padova
Via Giustiniani, 2
35100 Padova, Italy
E-mail: alex.pontini@libero.it

Summary

Purpose. Experimental animal study to evaluate the osseo integration, inflammatory response, dislocation and the reabsorption timing of the reabsorbable Mitek® mini anchor, in comparison with the non reabsorbable titanium based.

Methods. Twenty rabbits were used: divided into two groups of ten. Each animal underwent sectioning and reinsertion of the Achilles tendon bilaterally, using a reabsorbable mini anchor on one side and a controlateral non reabsorbable mini anchor. The first group was sacrificed after 40 days and the other after 120 days. The bone tendon complex was subjected to histological study.

Results. No histological and statistical significant difference were observed in each group, with a lower inflammation response in the reabsorbable implant. Electronic microscopy evaluation demonstrates good stability of the implant in each group. At day 120 the reabsorbable anchors were not yet disappeared.

Conclusion. Reabsorbable Mitek mini anchors are to be considered to guarantee a similar response and similar bone stability than non reabsorbable ones.

KEY WORDS: anchor, reabsorption, tendon, osseointegration, rabbit.

Introduction

The use of suture anchors has been copiously reported in the literature for the attachment of soft tissues to cortical and spongy bone¹. Orthopedic use of such suture include reconstruction of rotator cuff, Bankart's lesion's repair, meniscus tears or collateral ligament injuries treatment and different type of extremity injuries²⁻⁴. In the last few years their application has been expanded also to head and neck reconstruction and hand surgery, including various tendon injuries^{5,6}. Actually suture anchors are produced in a lot of different shapes, design and sizes, even reabsorbable (polylactic) than non reabsorbable (titanium) anchors. Different advantages of bioabsorbable implants in comparison with traditional metallic implants have been described^{7,8}. Reabsorbable implants seems to reduce stress shielding because their properties to gradually transfer load as they degrade. Depending of any specific application the bioabsorbable polymers can be also engineered to provide the desirable degradation profile.

Generally reabsorbable implant allows to reduce a second procedure for its removal and could facilitate a needed revision surgery. Use of foreign implants require also control done by magnetic resonance imaging (MRI) and bio absorbable implants offer a very low distortion rate.

First studies on bioabsorbable materials was published in 1966, when Kulkarni et al. published a report on the histological response and the degradation of poly-L-lactic acid (PLLA) powder implanted in guinea pigs and rats⁹. It was found that the polymer was nontoxic, nontissue reactive, and degraded slowly. The reabsorbable material of Mytek mini anchor is based on Poly-L-lactic acid (PLLA), that is the L isomer of polylactic acid. It is hydrophobic and crystalline and has an extended degradation time.

Traditional mitek mini anchors are composed of a titanium alloy shaft, represented by titanium (90%), aluminum (6%) and vanadium (4%). Their measure, even reabsorbable that not reabsorbable are 1.8 mm in diameter and 5.0 mm long.

Suture anchors design were thought to allow their placement into a bone hole obtained by a mechanical predrilling. Then the fixation of the anchor within the bone is provided by engaging their specific structures (arcs, barbs, or threads) attached to the anchor within the bone. A suture, generally reabsorbable, is available to be attached to the soft tissue, so they are directly fixed to the anchor. Suture anchors, have been

shown to be stronger than the attached sutures¹⁰. Fixation is more secure than periosteal suture because the soft tissue could be strongly attached to a bony embedded anchor. In hand surgery, the development in the last few years of the Mitek mini and micro suture anchor (Mitek Inc, Norwood, MA) to repair tendons, ligaments and the nearest structures, has permitted an important alternative to traditional pull-out technique, both transosseus¹¹ and subcutaneous, according to Mantero technique¹². However, the disadvantage of a permanent implant (titanium) could conduce to a secondary surgery for its removal in case of dislocation, fracture, infections or inflammatory response to a foreign body¹³. The reabsorbable implant in polylactic acid seems to avoid this complication¹⁴ but its utilisation in hand surgery is actually impeded and debated because of the poor scientific literature. The aim of this experimental study is to histological evaluate osseointegration properties of the two different types of Mitek mini anchors at different phases of healing.

Materials and methods

This study was approved by the Research Ethics Committees of the National Ministry of Health, and it was conducted in accordance with the principles of the Declaration of Helsinki regarding experimentation involving animals, in our Experimental Surgery Centre. The sample consisted of 20 male rabbits (*Oryctolagus cuniculus*) of the New Zealand White breed, weighing between 2200 and 3300 g. The animals were kept in the vivarium of our Experimental Surgery Centre, housed in adequately metal cages, with necessary hydration and food and weekly controlled from Veterinary Doctors of our University. The animals were divided randomly into two groups composed of 10 animals each: group A (40 days of survival) and group B (120 days of survival). Anaesthesia was administered with an intra muscle injection of 0,4 ml/kg of tiletamine chloridhydrate and zolazepam in

the proximal region of pelvic limbs, joint to a local injection of lidocaine, applied to the incision site. Antibiotic prophylaxis was administered, consisting of ofloxacin at dose of 5 mg/kg intramuscularly at the time of inducing anesthesia, and this was maintained for 6 days. Analgesia during the postoperative period, was obtained by 3 mg/kg of ketorolac, maintaining for 48-72 hours. After the onset of anesthesia, all the animals were subjected to surgical procedure. The area was accurately shaved with exposure of the skin in the calcaneus region and disinfected with povidone-iodine solution. An "S" incision was performed to expose the Achilles tendon and then the tendons of both pelvic limbs were sectioned adjacent to their insertions into the bone¹⁵, reproducing a tendon injury with the loss of bone insertion (Fig. 1). The sectioned tendons were then reinserted by using two different types of Mitek mini anchor, one side randomly decided with a non reabsorbable one, and the other side with a reabsorbable one. Using a 1.3 mm diameter drill, 2 holes were made into the calcaneus bone: the anchor was then fixed into it and with the non reabsorbable suture joint to the anchor was performed a tendon repair according to Kessler technique with a good fixation of the Achilles tendon to the calcaneus bone (Fig. 2). Both techniques were performed during the same operation on each rabbit, one on each pelvic limb. Thus, the same number of reinsertion was obtained for each mini anchor on the right and on the left side. The skin was then closed using nylon 5/0 thread and the operative wound was covered with a sterilized gauze compress. These 20 animals were divided into 2 groups with different postoperative periods and sacrificed afterwards: group A, 40 days; group B, 120 days. Following drug-induced sacrifice, the respective Achilles tendons were identified and isolated¹⁶.

The specimens of the pelvic limb, composed of a part of calcaneus bone containing the anchor and the tendon fixed to the bone were then extracted (Fig. 3). These were draped in dampened gauze with a saline solution of 0,9% sodium chloride and then the gauze was removed and the specimens were immediately placed in 10% neutral-buffered formalin. The speci-

Figure 1. In each animal an incision exposed the Achilles tendon and it was sectioned adjacent to insertion into the bone.

Figure 2. The anchor of each type was fixed in calcaneus bone and a tendon repair was performed.

Figure 3. The specimens, composed of a part of bone containing the anchor and the tendon fixed to the bone were then extracted.

mens were subsequently sent to our Institute of Pathology. The undecalcified calcaneus bones were dehydrated via ascending alcohol concentrations, and then they were impregnated with a 1:1 mixture of alcohol/Technovit 7200 VLC, followed by infiltration with pure Technovit 7200 VLC (Heraeus Kulzer, Germany). According to the Säge-Schliff (sawing and grinding) technique¹⁷ the embedded specimens were sawed along the longinal axis of the mini-Anchors with an Exakt® sawing machine to a thickness of about 100 micrometers and then ground with an Exakt® grinding machine to a thickness of 10 micrometers. The polished sections were stained with Haematoxylin-Eosin (H&E) and Masson-Goldner staining. Histological evaluation was performed on a Leica DM 4000B microscope, equipped with a Leica DFC 320 electronic camera. Each specimen was evaluated for intervening fibrous connective tissue, presence or absence of Osseo integration, anchor particulation and fracture, inflammatory cell infiltration, in comparison with non reasorbable titanium anchors and reasorbable polylactic anchors.

Results

At the time of extraction, each suture from the anchor to the tendon, in both groups was clinically intact, and each anchor was firmly embedded in the bone, except in one case (group B) in which both anchors were dislocated, probably because of the movement of the animal that was referred during the whole post-operative period. Light microscopic evaluation revealed in Group A –with no significant difference between reasorbable and non reasorbable anchors, –that the anchor was circumscribed with the bone, the eyelet end of each suture anchor was always deep to the cortical surface of the bone, no inflammatory cell infiltrate or foreign body giant cell reactions were evident. No particulation or fracture was associated with any of the anchors (Figs. 4, 5). In group B, the microscopic evaluation demonstrates

Figure 4. Group A. Titanium mini anchor implant at day 40. No difference with reasorbable implant were observed.

Figure 5. Group A. Polylactic mini anchor implant at day 40.

that titanium implants were good embedded in bone with no dislocation (Fig. 6). The reasorbable anchors had not disappeared, including extensive regions of direct mineral apposition (Fig. 7). Viable osteocytes were present near the anchor surface (Fig. 8). No difference in the amount of bone anchor apposition was noted among specimens. Inflammatory cell infiltrate and foreign giant cell reactions were evident in three specimens containing titanium anchor and were not demonstrated in specimens with polylactic anchors. Also in group B no particulation or fracture with any of the anchors was revealed. In both groups there were regions of direct bone contact on all surfaces of the anchors.

The mechanical assessment of the anchor function were performed by observation of rabbit movement and response to external stimulation done by veterinary doctors, where no tendon repair failure were noticed among groups. From day 2 after surgery till the period of sacrifice the rabbits demonstrate a normal capability of movement and jump into their cages, free rest positions and good capability of alimentation. Statistical analysis (T-Test) was performed considering medium value, in both group, at day 40 and 120, about inflammations cell count (number of inflammatory cell/mm²), osteocyte count (number of histological evident osteocyte/mm²), mineral apposi-

Figure 6. Group A. Particular of Titanium mini anchor implant at day 120. Implant was good embedded in bone.

Figure 7. Group B. Polilactic mini anchor implant at day 120. The reabsorption is not present yet. No inflammation were observed (particular).

Figure 8. Group B. Polilactic mini anchor implant at day 120. Viable osteocytes were present near the anchor surface suggesting initials Osseointegration process.

tion on anchor site (density range where evident at microscope). Significant difference was detected about inflammations ($p < 0,05$) demonstrating less foreign body response in reabsorbable groups. No significant difference ($p < 0,05$) were reported among groups in terms of osteocyte counts and mineral apposition (Tabs. 1-3).

Discussion

The tendon injuries represent an important challenge for hand surgery, particularly when it is not possible to perform a simple tendon suture but it is necessary to reinsert an avulse tendon. Therefore, many times, different techniques were created and evaluated for tendon fixation to the bone and with the help of technology by using an implant, especially miniaturized, like mini and micro suture anchor. The development of this technology conducted from a non reabsorbable implant (titanium) –that could have to be removed from the permanent implant, because of dislocation, infection or tissue reaction –to a reabsorbable implant (polylactic acid), that could avoid all these complications. In spite of the properties and the good results offered by the reabsorbable material^{18,19}, there was not much scientific evidence about the results that could be obtained with its use in hand surgery in terms of Osseo integration, inflammatory response, fracture, in comparison with titanium anchors that are still widely more used than non reabsorbable ones, as we have done in our Hand Surgery Unit. This experimental study reproduced in an animal model a typical tendon injury and a repair using bone fixation by mini anchor. Histological evaluation showed that reabsorbable Mitek mini anchor Osseo integrates in different periods of healing, as well as non reabsorbable Mitek mini anchor, with less inflammatory response in a long period of time. We also observed that after 120 days non reabsorbable anchors were still present, suggesting that their reabsorption is still longer as declared by the Mitek Company, and it demonstrates that we could speak about a slow reabsorption anchor as reported in other studies.

Animal studies using LPLA implants have demonstrated degradation time to be at least 3 years, that was a not applicable period in our animal model for evaluation. Resorption of polymers generally occurs in two phases²⁰. In the first phase, the polymer chains are broken down through hydrolysis. In this phase, the molecular weight drops first, followed by mechanical strength loss, and finally by a loss of mass²¹. In the second phase, the implant loses its form and breaks physically into particles, which are attacked by macrophages. Depending on the size of the particulates, they are phagocytosed and the byproducts are excreted by the kidneys and lungs. The corresponding biological response to the degrading polymer is thought to happen as a result of either a build up of acidic degradation products or as a response to the particulates of the polymer²². The tim-

Table 1. Inflammatory cells counts among two type of implants in rabbits sacrificed at day 40 and 120 (medium value).

Rabbits on day 40 of sacrifice / on day 120	Tendon with reasorbeable material; inflammatory cell/counts – mm ²	Tendon with titanium implants; inflammatory cell/counts – mm ²
Rabbit n.1	19 / 46	17/34
Rabbit n.2	16/25	19/27
Rabbit n.3	23/48	25/45
Rabbit n.4	27/112	24/110
Rabbit n.5	45/115	39/117
Rabbit n.6	6/24	8/23
Rabbit n.7	9/14	7/13
Rabbit n.8	18/34	17/38
Rabbit n.9	16/32	15/37
Rabbit n.10	24/87	25/88

Table 2. Viable osteocyte counts among two type of implants in rabbits sacrificed at day 40 and 120 (medium value).

Rabbits on day 40 of sacrifice / on day 120	Tendon with reasorbeable material; Viable osteocyte– mm ²	Tendon with titanium implants; viable osteocyte – mm ²
Rabbit n.1	29/67	19/68
Rabbit n.2	25/113	23/118
Rabbit n.3	28/115	29/117
Rabbit n.4	28/116	26/119
Rabbit n.5	45/78	47/81
Rabbit n.6	14/64	16/61
Rabbit n.7	18/65	20/70
Rabbit n.8	56/119	57/121
Rabbit n.9	18/45	19/49
Rabbit n.10	25/58	25/56

Table 3. Mineral apposition (mm) among two type of implants in rabbits sacrificed at day 40 and 120 (medium value).

Rabbits on day 40 of sacrifice / on day 120	Tendon with reasorbeable material; mm mineral apposition	Tendon with titanium implants; mm mineral apposition
Rabbit n.1	0.04/0.07	0.05/0.06
Rabbit n.2	0.03/0.08	0.04/0.09
Rabbit n.3	0.01/0.03	0.03/0.04
Rabbit n.4	0.04/0.06	0.03/0.06
Rabbit n.5	0.06/0.08	0.06/0.07
Rabbit n.6	0.04/0.05	0.05/0.07
Rabbit n.7	0.05/0.09	0.05/0.08
Rabbit n.8	0.04/0.06	0.03/0.06
Rabbit n.9	0.05/0.09	0.04/0.07
Rabbit n.10	0.03/0.09	0.04/0.08

ing of the foreign-body response is thought to be related to the final stage of polymer degradation, so it explain as in our histological findings it is not represented. Anyway, according to our clinical opinion, a long time of reabsorption is more useful to permit a better Osseo integration and a good stress resistance, offering the opportunity for a better and more stable tendon healing, even also with low inflammatory response that could optimize tendon repair. Experimental studies cannot provide conclusions that are directly applicable to making decisions about the choice of material for the application to humans injuries. Nevertheless, tendon reinsertion by suture anchor is simpler and was demonstrated to be efficient in tendon avulsion and the use of a reasorbeable material could be considered similar to a non reasorbeable one with, moreover, a less inflammatory response in the period we have considered.

It could be interesting to evaluate, in a larger period of time, the reabsorption modality and how the new bone, that we observe in its first time of growth, could substitute the polylactic anchor. At this time of our study we could suggest that reasorbeable Mitek mini anchors offered the same security of non reasorbeable ones in terms of Osseo integration, stable fixation and, at the same time, avoided the possible complications linked to the use of a permanent implant. Besides we encouraging further experimental or clinical trial that could support our *in vivo* and histological observation.

References

- Goble EM, Somers WK, Clark R, Olsen RE. The development of suture anchors for use in soft tissue fixation to bone. *Am J Sports Med* 1994 Mar-Apr; 22(2):236-239.

2. Bhagia SM, Ali MS. Bankart operation for recurrent anterior dislocation of the shoulder using suture anchor. *Orthopedics* 2000; 23(6):589-591.
3. Lynch SA, Beard DM, Renström PA. Repair of distal biceps tendon rupture with suture anchors. *Knee Surg Sports Traumatol Arthrosc* 1999; 7(2):125- 131.
4. McDermott TP, Levin LS. Suture anchor repair of chronic radial ligament injuries of the metacarpophalangeal joint of the thumb. *J Hand Surg* 1998; 23B(2):271- 274.
5. Dzwierzynski WW, Sanger JR, Larson DL. Use of Mitek suture anchors in head and neck reconstruction. *Ann Plast Surg* 1997; 38(5):449-454.
6. Francis KR, Hoffman LA, Cornell C, Cortese A. The use of Mitek anchors to secure mesh in abdominal wall reconstruction. *Plast Reconstr Surg* 1994; 93(2):419-421.
7. Südkamp NP, Kääb MJ. Biodegradable implants in soft tissue refixation: experimental evaluation, clinical experience, and future needs. *Injury* 2002; 33(Suppl 2).
8. Ambrose CG, Clanton TO. Bioabsorbable implants: review of clinical experience in orthopedic surgery. *Ann Biomed Eng* 2004; 32:171-177.
9. Kulkarni RK, Pani KC, Neuman C, Leonard F. Polyactic acid for surgical implants. *Arch Surg* 1966; 93:839-843.
10. Baeber FA, Herbert MA, Richards DP. Sutures And suture anchors : update 2003. *Arthroscopy* 2003; 9:109-113.
11. Bunnel S. Gig pull-out suture for tendons. *J Bone Joint Surg* 1954; 36A:850-851.
12. Guinard D, Montanier F, Thomas D, Corcella D, Moutet F. The Mantero flexor tendon repair in zone 1. *J Hand Surg* 1999; 24B:755-757.
13. Giannikas D, Athanaselis E, Matzaroglou C, Saridis A, Tyl-
lianakis M. An unusual complication of Mitek suture anchor use in primary treatment of flexor digitorum profundus tendon laceration: a case report. *Cases J* 2009; 2:9319.
14. Claes L, Ignatius A. Development of new, biodegradable implants. *Chirurg* 2002; 73:990-996.
15. Shaieb MD, Singer DI, Grimes J, Namiki H. Evaluation of tendon -to-bone re attachment: a rabbit model. *Am J Orthop* 2000; 29:537-542.
16. Matos MA, Tannuri U, Guarniero R. Effect of ifosfamide on bone healing. *J Musculoskeletal Neuronal Interact* 2006; 6:247-250.
17. Donath K, Breuner G. A method for the study of undecalcified bones and teeth with attached soft tissues. The Säge-Schliff (sawing and grinding) technique *J Oral Pathol* 1982; 11:318-326.
18. Madhavan Nampoothiri K, Nair NR, John RP. An overview of the recent developments in polylactide (PLA) research. *Bioresour Technol* 2010; 101:8493-8501.
19. Kontakis GM, Pagkalos JE, Tosounidis TI, Melissas J, Kantonis P. Bioabsorbable materials in orthopaedics. *Acta Orthop Belg* 2007; 73(2):159-169.
20. Kohn J, Langer R. Bioresorbable and bioerodable materials. In: *Biomaterials Science: An Introduction to Materials in Medicine*, edited by B. D. Ratner, A. S. Hoffman, F. J. Schoen and J. E. Lemons. San Diego, CA: Elsevier 1996; 64-72.
21. Pietrzak WS, Sarver D, Verstynen M. Bioresorbable implants-practical considerations. *Bone* 1996; 109S-119.
22. Bergsma JC, De Bruijn WC, Rozema FR, Bos RRM, Boering G. Late degradation tissue response to poly (l-lactide) bone plates and screws. *Biomaterials* 1995; 16:25-31.